



# Anhui Intco Medical Products Co., Ltd

Address: Haitang road west and yinhua road north) Suixi Wuhu  
 Modern Industrial Park, Suixi Town, Huaibei City, Anhui  
 Province, P. R. China.

Tel:++86 21 34978818 Fax:+86 21 34978808

Web:www.intcomedical.com

## Examination Disposable Gloves Product Specification

Type	Disposable Nitrile Gloves, Powder free, Non Sterile,Ambidextrous														
Grade	Examination		Surface	Finger texture			Glove Length(mm)	≥240							
Material	NBR		Packing	100pcs/box,10boxes/ctn			Weight(g)	3.5±0.3 (Medium Size)							
Cuff	Beaded					Powder Level	≤2mg/glove								
Product	Color	Size	Weight	cal Dimension					Physical Property				Water Test AQL	Appearance AQL	
				Length (mm)	Palm Width (mm)	Cuff Thickness(mm)	Finger Thickness(mm)	Palm Thickness(mm)	Tensile Strength(N)		Elongation				
										Before Aging	After Aging	Before Aging	After Aging		
Nitrile gloves	Blue/White/Purple	S	3.0g±0.3	≥240	85±5	0.07±0.03	0.10±0.03	0.08±0.03	6.0	6.0	500	400	AQL1.5	Critical AQL1.5 Major AQL2.5 Minor 4.0	
		M	3.5g±0.3		95±5				6.0	6.0	500	400	AQL1.5	Critical AQL1.5 Major AQL2.5 Minor 4.0	
		L	3.8g±0.3		110±5				6.0	6.0	500	400	AQL1.5	Critical AQL1.5 Major AQL2.5 Minor 4.0	
		XL	4.1g±0.3		120±5				6.0	6.0	500	400	AQL1.5	Critical AQL1.5 Major AQL2.5 Minor 4.0	
No.	sort		Critical defects				Major defects				Minor defects				
1	torn	big tearing	√				-				-				
		scratch	≥3mm				≥2mm				-				
		visible hole	≥3mm				≥3mm				-				
2	dirt	greasy dirt	≥3mm,serious				1-2mm,visible				<2mm,slightly				
		yellow dirt	√				-				-				
3	black spots,impurities		uncountable				≥3mm or 0.5-3mm,>3/side				0.5-3mm,≤3/side				
4	beading		broken;torn;none-beading				impurity;knot;black;bright PU≥3mm;inward≥10mm;minuteness(≤0.5cm)				inward<10mm;thin 0.5-0.8				
5	sticky		-				sticky cuff severely				slightly				
6	Extreme color variation		entirely,severely				finger,cuff,severely				cuff,slightly,unconspicuous				
7	Bad smell		-				severely irritating smell				slight odor				
8	Embedded foreign particles		insect,hair,etc.				-				-				
We has established the Quality management system in according with ISO9001 and ISO 13485 standards															
The certification which we have is EN455,EN374 ,CE certification ,ISO9001 ISO13485															
Sampling Procedures according to the standard of ISO2859,EN455, ASTM D6319,Inspection Level for physical dimensions and force /elongation at break is according to S-2 ,water test according to G-1															